

# SP-600 Quick Start Guide

## Conductivity Testing



1) Press **OK** to power-up. The Conductivity test screen is the Main screen and is ready for testing.



2) Insert and triple rinse sample water in main cell. When you see the blue bar it has locked in the value.

## Cleaning Kit for Pyxis Handheld Devices



Pyxis recommends a minimum cleaning frequency of once per month be maintained dependent on application needs and foulant level. High stress applications with excessive suspended solids and corrosion/scale by-product can result in the need to increase the frequency of cleaning your handheld device.



Scan QR code to watch a demonstration on how to cleaning your Pyxis meter

For information on ordering the Pyxis Handheld Meter Cleaning Solution, Scan the QR code to take you to the Pyxis website.



## Chlorine Testing



1) Press **< Cond** to select Chlorine, click **OK** button to start testing.

2) Select Method, press **OK** button.

3) Triple rinse sample water, press **Zero <**. Remove sample from cell.



3) . Add DPD powder Pillow to 10 ml vial with the sample water. Mix well. Transfer the developed sample to sample cell.



4) Press **>/Timer** to start 1 minute timer.



5) Press **OK/ Read** button for results

# Pyxis®

## SP-600 pH/ORP Module



### Bluetooth Wireless

- Fully Sealed & Separated Module
- Wireless Data Transfer to SP-600
- Interfaces with uPyxis App & SP-910
- Locking firmware



When the pH Module is turned on, it will stay on for the duration of time when the body main body is on. It has a permanently sealed 3.7 V lithium battery

The pH/ORP module will turn itself off when either commanded by the main SP-600 module or automatically based on the idle time limit (180 seconds).

The time setting is customizable via the uPyxis Mobile app or the uPyxis desktop app.

pH Modules have a 6-month warranty with the typical life span 12-18 months. When not in use, keep Pyxis pH/ORP Sensor Storage Solution in module.



Scan QR Code for more information or ordering a replacement module.

Please maintain 2 ml of the Pyxis pH storage solution in the pH/ORP cell at all times to keep the electrodes hydrated.

**Need Support! Email or Call us by using the QR Codes below**



Scan QR Code to send email to the Pyxis Service Department

Scan QR Code to make a phone call to the Pyxis Service Department





# SP-600 Quick Start Guide



pH/ORP Sensor  
Storage Solution

SKU: 63900



pH Reference Standard  
Combo Kit

SKU: 57007



Conductivity Reference  
Standard 1000 µS

SKU: 57008



DPD Chlorine Secondary  
Standard 1 ppm

SKU: 21039



Free Chlorine Powder Pillow  
0.00 - 2.2 mg/L 100 pk  
10 ml test

SKU: 57002



Total Chlorine Powder Pillow  
0.00 - 2.2 mg/L 100 pk  
10 ml test

SKU: 58-1



Scan QR Code to order on:  
<http://www.pyxis-lab.com>

## pH Testing



1) Press > to select pH mode to connect  
mainbody to pH/ORP module.



2) Once connected you will see the prompt "Cell Empty?",  
flip open the lid and triple rise the pH/ORp solution out of cell.

3) Triple rinse the sample water in pH/ORP cell for reading.



4) The blue bar will show the pH value is locked in.

Please maintain 2 ml of the Pyxis pH/ORP storage solution in the  
pH/ORP cell at all times to keep the electrodes hydrated.

# Pyxis®

## SP-600 Multi Meter Quick Start Guide



[Quick Link to SP-600  
Operation Manual](#)



SP-600 Meter

Scan the QR code for SP-600  
Operation Manual

[www.pyxis-lab.com](http://www.pyxis-lab.com)

Pyxis Lab, Inc  
1729 Majestic Dr., Ste 5  
Lafayette, CO. 80026  
866-203-8397  
[service@pyxis-lab.com](mailto:service@pyxis-lab.com)